

TXL-GEN-009 / 27 October 2026 · One-year manufacturer warranty

Conductivity Meter

Solution conductivity measuring instrument. A functional device that meets the basic needs of every textile laboratory, offering regular and flexible operation capabilities.



Model technical data

1	RS232 interface allows for connection to external devices for data storage or transfer.
---	---

Physical specifications and utilities

Electrical supply	Approx. low-voltage adapter with 100–240 V AC, 50/60 Hz input; output voltage is build-dependent.
Power / connected load	Approx. 5–20 W connected load for preliminary planning.
Overall instrument dimensions	Approx. 250 × 250 × 120 mm (W × D × H).
Instrument weight	Approx. 0.8–2 kg net.
Gross / shipping weight	Approx. 2.9–4.5 kg gross; estimated packing allowance.
Interfaces	RS232 interface allows for connection to external devices for data storage or transfer.
Lab requirements	Level, stable working surface; specimen conditioning follows the selected method.

Dimension order and approximate values are stated in each row. Packaging data are separate from instrument dimensions and net mass. Approx. rows are preliminary engineering estimates based on comparable equipment and function/capacity; they are not verified manufacturing measurements or standard requirements. Installation and shipment are finalised against the approved drawing and nameplate. Estimate scope: base product / one unit; external computer, bench and auxiliary plant excluded.

Specifications, dimensions, and power requirements are approximate and subject to change without prior notice for continuous design improvements.

Product technical specifications and images may vary according to manufacturing conditions and applicable standards.

Test and configuration

Measured output / function	Electrical conductivity ($\mu\text{S}/\text{cm}$ or mS/cm)
Parameters selected for the method	Cell constant, temperature and sample range
Verification points	Verification using a traceable conductivity standard

Control and recording specification

Method-specific software and recording requirements follow. Delivered software modules are defined by the order scope.

- Method code and edition, specimen ID, operator and selected test parameters are linked to the record.
- The calculation and endpoint are defined by the selected method edition; instrument signals and operator assessments are kept distinct.
- Raw measurements, units, valid/invalid test flags and the reported result remain traceable.
- Calculations are checked against known inputs and an independent calculation; sensor channels are checked against calibration values.

Related test methods

-

Listed methods are application references. Suitability is established by assessing hardware, method edition and calibration together.