

TXL-KMS-037 / 27 October 2026 · One-year manufacturer warranty

Sewability Tester

Tests fabric sewability independently of sewing conditions or operator skill. It measures needle penetration force; the value is read in cN and can be converted to other units. USB data output is standard; PC software and printer are optional.



Model technical data

1	Tests fabric sewability independently of sewing conditions or operator skill.
2	Needle penetration force measurement is performed.
3	The value is read in cN and can be converted to different units of measurement.
4	USB data output is standard; PC software and printer are optional.

Physical specifications and utilities

Electrical supply	Approx. 220–240 V AC, 50/60 Hz, single phase; build-dependent.
Power / connected load	Approx. 100–400 W connected load for preliminary planning.
Overall instrument dimensions	Approx. 700 × 450 × 600 mm (W × D × H).
Instrument weight	Approx. 25–50 kg net.
Gross / shipping weight	Approx. 37.5–72.5 kg gross; estimated packing allowance.
Interfaces	USB data output is standard; PC software and printer are optional.
Lab requirements	Level, stable working surface; specimen conditioning follows the selected method.

Dimension order and approximate values are stated in each row. Packaging data are separate from instrument dimensions and net mass. Approx. rows are preliminary engineering estimates based on comparable equipment and function/capacity; they are not verified manufacturing measurements or standard requirements. Installation and shipment are finalised against the approved drawing and nameplate. Estimate scope: base product / one unit; external computer, bench and auxiliary plant excluded.

Specifications, dimensions, and power requirements are approximate and subject to change without prior notice for continuous design improvements.

Product technical specifications and images may vary according to manufacturing conditions and applicable standards.

Test and configuration

Measured output / function	Needle penetration force and sewability
Parameters selected for the method	Needle type, specimen structure, speed and cycles
Verification points	Force sensor, needle and travel speed

Control and recording specification

Method-specific software and recording requirements follow. Delivered software modules are defined by the order scope.

- Method code and edition, specimen ID, operator and selected test parameters are linked to the record.
- The calculation and endpoint are defined by the selected method edition; instrument signals and operator assessments are kept distinct.
- Raw measurements, units, valid/invalid test flags and the reported result remain traceable.
- Calculations are checked against known inputs and an independent calculation; sensor channels are checked against calibration values.

Related test methods

-

Listed methods are application references. Suitability is established by assessing hardware, method edition and calibration together.